

Installation & User Instructions for the Solvent Sentry™ Level Sensor

1

Attach the Solvent Sentry to an Argento Scientific or third party manifold cap.

2

Pair the Solvent Sentry to the hub by lifting the float on the sensor and holding down the "pair" button on the hub.

3

Begin running chromatography as normal.

4

When the alarm sounds and the light flashes, safely empty the waste container.

5

Reattach cap to waste container. The level sensor automatically resets.

sensor housing

manifold

float (inside view)

hub

